

Auditdata

DECLARATION OF CONFORMITY

REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT

Manufacturer name and address	Auditdata A/S Wildersgade 10B 1408 Copenhagen Denmark	
Notified Body name and address	TÜV SÜD Product Service GmbH Ridlerstrasse 65 80339 München	 0123
Product Identification	MD Category: Trademark: Type/Model: CS (Common specification) SRN: Basic UDI/DI: Risk class: Lot/Batches/Serial number:	Hearing Medical Diagnostic (Hardware and Software) Primus PFU+ Primus Fitting Unit - PFU+ (PFU+) N/A no common specification has been published DK-MF-000011415 05711781DHF2000ZC IIa, rule 10 All issued serial numbers from 21000001
Intended purpose	Audiometer is a device used for evaluating hearing acuity. The audiometer records the subject's responses to produce an audiogram of threshold sensitivity, or speech understanding profile. Audiometer with stated accessories is indicated for non-continuous, noninvasive air and optionally bone conduction and speech audiometric testing. Audiometer is indicated for non-continuous real-ear measurements (REM) at the ear drum by means of noninvasive external ear canal insertion of a probe tube.	
Conformity assessment	Annex IX (Quality system and technical documentation assessment)	
EC-Certificate No.:	G10 076081 0015	
DOC valid until	2029-02-18	

This declaration of conformity is issued under the sole responsibility of Auditdata A/S. We hereby declare that the medical device specified above is in conformity with the European Regulation (EU) 2017/745 and Directive 2011/65/EU.

Copenhagen, February 19th 2024

Denys Lebedev, QA/RA Manager

