## **Auditdata**

## DECLARATION OF CONFORMITY

## REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT

Manufacturer name and

address

Auditdata A/S Wildersgade 10B 1408 Copenhagen

Denmark

Notified Body name and

address

TÜV SÜD Product Service

Gmbh

Ridlerstrasse 65 80339 München  $(\epsilon)$ 

**Product Identification** MD Category: Hearing Medical Diagnostic

(Hardware)

Trademark: Entomed Audiometer

Type/Model: SA 202

CS (Common specification) N/A no common specification

has been published

SRN: DK-MF-000011415
Basic UDI-DI: 05711781DHF2009ZW

Risk class: Ila, rule 10

Lot/Batches/Serial number: All issued serial numbers from

EE152001

**Intended purpose** Audiometer is a device used for evaluating hearing acuity. The

audiometer records the subject's responses to produce

an audiogram of threshold sensitivity, or speech understanding

profile.

**Conformity assessment** Annex IX (Quality system and technical documentation

assessment)

**EC-Certificate No.:** G10 076081 0015

**DOC valid until** 2029-02-18

This declaration of conformity is issued under the sole responsibility of Auditdata A/S. We hereby declare that the medical device specified above Is in conformity with the European Regulation (EU) 2017/745 and Directive 2011/65/EU.

Copenhagen, February 20th 2024

Denys Lebedev, QA/RA Manager