

# Auditdata

## DECLARATION OF CONFORMITY

REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT

**Manufacturer name and address**

Auditdata A/S  
Wildersgade 10B  
1408 Copenhagen  
Denmark

**Notified Body name and address**

TÜV SÜD Product Service  
GmbH  
Ridlerstrasse 65  
80339 München



**Product Identification**

MD Category:	Hearing Medical Diagnostic (Hardware and Software)
Trademark:	Primus Pro
Type/Model:	Primus Fitting Unit Pro (2000 Primus Fitting Unit Pro, PFU Pro)
CS (Common specification)	N/A no common specification has been published
SRN:	DK-MF-000011415
Basic UDI/DI:	05711781DHF2000ZC
Risk class:	Ila, rule 10
Lot/Batches/Serial number:	All issued serial numbers from 25000001

**Intended purpose**

Audiometer is a device used for evaluating hearing acuity. The audiometer records the subject's responses to produce an audiogram of threshold sensitivity, or speech understanding profile. Audiometer with stated accessories is indicated for non-continuous, noninvasive air and optionally bone conduction and speech audiometric testing. Audiometer is indicated for non-continuous real-ear measurements (REM) at the ear drum by means of noninvasive external ear canal insertion of a probe tube.

**Conformity assessment**

Annex IX (Quality system and technical documentation assessment)

**EC-Certificate No.:**

G10 076081 0015

**DOC valid until**

2029-02-18

This declaration of conformity is issued under the sole responsibility of Auditdata A/S. We hereby declare that the medical device specified above is in conformity with the European Regulation (EU) 2017/745 and Directive 2011/65/EU.

**Copenhagen, February 19<sup>th</sup> 2024**

Denys Lebedev, QA/RA Manager